

Government of West Bengal
Office of the Chief Medical Officer of Health
District Health and Family Welfare Samiti, Cooch Behar
Lalbagh: Cooch Behar
Tel: 228874(03582) Fax: 228966
E-mail: cmoh.cbr@gmail.com

Memo No : 1947

Dated :- 21/4/22

QUOTATION NOTICE (2nd Call)
Ref NIQ No:-1399 Dated-22/03/2022

Sealed quotation is invited from reputed firm /agency for supply Carbon Monoxide Breath Monitor. Interested parties are requested to submit their quotation in sealed cover addressed to the Chief Medical Officer of Health & Secretly, DH & FW Samiti, Cooch Behar Superscribing "**Quotation for supply Carbon Monoxide Breath Monitor**".

Detailed specification of is mentioned below:-


Sl No	Name of the items	Specification	Quantity
1	Carbon Monoxide Breath Monitor	Attached as Annex-A	01 Pcs

Last date of submission of quotation is 28/04/2022 before 5.30 PM. Photocopy of trade license, last income tax return and PAN card & GST No must be submitted with the quotation document. Quotation will be opened on 29/04/2022 at 2.30 pm in the chamber of the undersigned.


CMOH & Secretary

Dist. H.& F. W. Samiti ,Cooch Behar

Memo No:-


21/4/22

Dated :-

Copy forwarded for information and with the request to hanging the same in the notice Board;-

1. The Sabhadhipati, Cooch Behar Zilla Parishad, Cooch Behar
2. The District Magistrate, Office of the District Magistrate, Cooch Behar
3. Dy. CMOH-I, Cooch Behar
4. Dy. CMOH-II, Cooch Behar
5. Dy. CMOH-III, Cooch Behar
6. Officer In Charge (Health) , DM Office, Cooch Behar
7. Accounts Officer, CMOH Office, Cooch Behar
8. District Information and Cultural Affairs Officer, Cooch Behar
9. The Station Manager, Railway Station, Cooch Behar
10. The Post Master, Head Post Office, Cooch Behar
11. The SMVP, MJNMCH, Cooch Behar
12. Officer In Charge, NIC, Cooch Behar with request to upload the quotation notice in website [www. coochbehar.gov.in](http://www.coochbehar.gov.in) of Cooch Behar
13. Dispatch section of this Office to display the notice in the office Notice Board.

CMOH & Secretary
Dist. H.& F. W. Samiti, Cooch Behar

Annex - A

Technical specifications Essential

Specifications:

1. It should have color touch-screen display.
2. Response time should be < 30 seconds to 90% FSD
3. Recording and interpreting results should be quick and easy
4. It should have automatic calibration to ensure accurate results
5. There should be provision of mouthpieces for excellent and low-cost infection control
6. It should have storage of up to 100 readings and personal profiles
7. There should be familiar green, amber and red traffic light system for making CO levels instantly identifiable to patients.
8. It should have electrochemical sensor with $\pm 5\%$ repeatability and accuracy.
9. Sensor operating life should be 5 years (2-year warranty)
10. Sensor sensitivity should be 1ppm.
11. Its weight should be in < 250 gm.
12. H₂ cross sensitivity: $< 12\%$

Desirable Specifications:

1. It should be able to provide instant result in exact ppm for %COHb and %FCOHb
2. There should be provision for adults, adolescents and pregnant woman testing mode

Environmental factors

1. Operation temperature range should be 0-40 degree Celsius
2. Storage/transport temperature: 0-50°C
3. Operating/storage/transport pressure: Atmospheric $\pm 10\%$
4. Operating humidity: 15-90% non-condensing
5. Storage/transport humidity: 0-95%

Documentation

QUOTATION DOCUMENT

**TECHNICAL AND FINANCIAL BIDS ON ACCOUNT OF THE MEMO NO:- 1947
DATED-21/04/2022 FOR CARBON MONOXIDE BREATH MONITOR UNDER NUHM**

THE CMOH OFFICE, COOCH BEHAR

SL No	Name of the items	Spe-dfication	Rate quoted inclusive of all taxes	Time required to supply after getting supply order (not more than 15 day)
I.	Carbon Monoxide Breath Monitor	Annex-A		

Date :-.....

Place :-.....

Sign of bidder with date or seal (if.....)